

Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-2626
Production date	15 July 2026
Expiry date	31 January 2027
Total volume (ml)	3000
Batchnumber HSA	B23K049493
Batchnumber HYA	S2015H40-300320
Product was sterilized by aseptic processing techniques	

Parameter	Results	Specifications
pH	7,49	7,30-7,60
Osmolality	280	270-290 mOsm/kg
Endotoxin	< 0,13	<0,5 EU/ml
Concentration HSA	4,1	3,2-4,8 g/l
Enzyme activity	52-76	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	30 July 2026
Released by	Dr. René Richarz
Signature	

REV 09_00 /2023-09-15