

Certificate of Analysis GM501 Hyaluronidase

Batchnumber	HYA-2635
Production date	30 September 2026
Expiry date	30 September 2027
Total volume (ml)	6000
Batchnumber HSA	B23K064693
Batchnumber HYA	S2015H40
Product was sterilized by aseptic processing techniques	

Parameter	Results	Specifications
pH	7,35	7,30-7,60
Osmolality	279	270-290 mOsm/kg
Endotoxin	< 0,16	<0,5 EU/ml
Concentration HSA	3,8	3,2-4,8 g/l
Enzyme activity	56-83	40-160 IU/ml
Sterility test (according to the current Ph.Eur.2.6.1 /USP<71>)	No growth	No growth
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	16 October 2026
Released by	Dr. René Richarz
Signature	

REV 09_00 /2023-09-15