

**Certificate of Analysis
GM501 Mineral Oil**

Batchnumber	OIL-1706
Production date	14 February 2017
Expiry date	31 August 2018
Total volume (ml)	1500

	Results	Specifications
Appearance	PASSED	PASSED
Sterility test Sterilised by sterile filtration over 0,2µm filter	PASSED	PASSED
Acidity	Conform Ph Eur	Conform Ph Eur
Solid paraffin	Conform Ph Eur	Conform Ph Eur
Viscosity	20,8	< 30 cP at 30°C
Specific gravity	0,8454	< 0,86 g/ml
Endotoxin	< 0,1	< 0,25 EU/ml
Peroxide value (POV)	< 0,1	< 0,1 mEq/kg
Mouse Embryo Assay (1-cell to expanded blastocyst after 120 hours)	95,24	≥ 80 %

Release date 06 March 2017

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann