

Certificate of Analysis GM501 Mineral Oil

Batchnumber	OIL-1712
Production date	13 April 2017
Expiry date	31 October 2018
Total volume (ml)	60000

	Results	Specifications
Appearance	PASSED	PASSED
Sterility test <small>Sterilised by sterile filtration over 0,2µm filter</small>	PASSED	PASSED
Acidity	Conform Ph Eur	Conform Ph Eur
Solid paraffin	Conform Ph Eur	Conform Ph Eur
Viscosity	20,2	< 30 cP at 30°C
Specific gravity	0,8453	< 0,86 g/ml
Endotoxin	< 0,1	< 0,25 EU/ml
Peroxide value (POV)	< 0,1	< 0,1 mEq/kg
Mouse Embryo Assay <small>(1-cell to expanded blastocyst after 120 hours)</small>	100	≥ 80 %

Release date 28 April 2017

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann