

**Certificate of Analysis
GM501 PVP**

Batchnumber	PVP-1703
Production date	28 February 2017
Expiry date	30 November 2017
Total volume (kits)	386
Batchnumber HSA	A4DF500171

	Results	Specifications
pH	7,49	7,20-7,60
Osmolality	293	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 1,00 EU/mL
Viscosity in cP (26°C)	217,2	> 170 cP
Mouse Embryo Assay (1-cell)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	17 March 2017
Released by	Vanessa Bierlich
Signature	

P.P. 