

## Certificate of Analysis GM501 PVP

Batchnumber	PVP-1708
Production date	18 May 2017
Expiry date	28 February 2018
Total volume (kits)	100
Batchnumber HSA	A4DF500171

	Results	Specifications
pH	7,29	7,20-7,60
Osmolality	304	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 1,00 EU/mL
Viscosity in cP (26°C)	195,6	> 170 cP
Mouse Embryo Assay (1-cell)	95	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 07 June 2017

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*