

**Certificate of Analysis
GM501 PVP**

Batchnumber	PVP-1811
Production date	19 June 2018
Expiry date	31 March 2019
Total volume (kits)	732
Batchnumber HSA	B3AKA00593

	Results	Specifications
pH	7,43	7,20-7,60
Osmolality	298	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,0705	< 1,00 EU/mL
Viscosity in cP (26°C)	254,3	> 170 cP
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 09 July 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann