

**Certificate of Analysis  
GM501 PVP**

Batchnumber	PVP-1912
Production date	13 June 2019
Expiry date	31 March 2020
Total volume (kits)	290
Batchnumber HSA	K824B6871

	<b>Results</b>	<b>Specifications</b>
pH	7,36	7,20-7,60
Osmolality	304	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 1,00 EU/mL
Viscosity in cP (26°C)	239,9	> 170 cP
Mouse Embryo Assay (1-cell)	80,95	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 28 June 2019

Released by Vanessa Bierlich

Signature