

Certificate of Analysis GM501 PVP

Batchnumber	PVP-2014
Production date	29 September 2020
Expiry date	30 June 2021
Total volume (ml)	750
Batchnumber HSA	B3ALD01343

	Results	Specifications
pH	7,46	7,20-7,60
Osmolality	288	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 1,00 EU/mL
Viscosity in cP (26°C)	257,9	> 170 cP
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	21 October 2020
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann