

**Certificate of Analysis  
GM501 PVP**

Batchnumber	PVP-2204
Production date	21 February 2022
Expiry date	30 November 2022
Total volume (ml)	640
Batchnumber HSA	B3ALF01143

	<b>Results</b>	<b>Specifications</b>
pH	7,45	7,20-7,60
Osmolality	305	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 1,00 EU/mL
Viscosity in cP (25°C)	234	> 170 cP
Mouse Embryo Assay (1-cell)	100	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date	09 March 2022
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*