

**Certificate of Analysis  
GM501 PVP**

Batchnumber PVP-2415  
 Production date 16 May 2024  
 Expiry date 28 February 2025  
 Total volume (ml) 782  
 Batchnumber HSA B03H100413

<b>Parameter</b>	<b>Results</b>	<b>Specifications</b>
pH	7,31	7,20-7,60
Osmolality	290	280-320 mOsm/kg
Viscosity (at 25°C)	170,4	> 170 cP
Endotoxin level	< 0,26	< 0,5 EU/mL
Concentration HSA	3,7	3,2-4,8 g/l
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
Mouse Embryo Assay (% blastocysts after 96 hours, exposure time to test medium:10 minutes)	100	≥ 80 %
HSSA (% motility compared with control after 60 minutes exposure to test medium)	90	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 03 June 2024  
 Released by Joana Sturm  
 Signature 

REV 08\_00 / 2023-09-15

