GYNEMED

# Certificate of Analysis
# GM501 PVP

| | |
|---|---|
| Batchnumber | PVP-2630 |
| Production date | 26 October 2026 |
| Expiry date | 31 July 2027 |
| Total volume (ml) | 720 |
| Batchnumber HSA | B03J125033 |

Product was sterilized by aseptic processing techniques

| Parameter | Results | Specifications |
|---|---|---|
| pH | 7,33 | 7,20-7,60 |
| Osmolality | 303 | 280-320 mOsm/kg |
| Viscosity (at 25°C) | 201,6 | > 170 cP |
| Endotoxin level | < 0,1 | < 0,5 EU/mL |
| Concentration HSA | 3,9 | 3,2-4,8 g/l |
| Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>) | No growth | No growth |
| Mouse Embryo Assay (% blastocysts after 96 hours, exposure time to test medium:10 minutes) | 100 | ≥ 80 % |
| HSSA (% motility compared with control after 60 minutes exposure to test medium) | 97 | ≥ 80 % |

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphillis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

| | |
|---|---|
| Release date | 12 November 2026 |
| Released by | Nicole Geiger |
| Signature | pp. N. Geiger |



Gynemed GmbH Co. KG, Wagrienring 24b, 23730 Sierksdorf, Germany
phone: +49-4561-5132-500, fax: +49-4561-5132-519, e-mail: info@gynemed.de