

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-1701
Production date	17 February 2017
Expiry date	31 August 2017
Total volume (mL)	970

	Results	Specifications
pH	7,29	7,20-7,60
Osmolality	306	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 45 min exposure	100	≥ 80 %

Release date 06 March 2017

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann