

**Certificate of Analysis  
GM501 SpermMobil**

Batchnumber	SMOBIL-1702
Production date	04 April 2017
Expiry date	31 October 2017
Total volume (mL)	1400

	<b>Results</b>	<b>Specifications</b>
pH	7,24	7,20-7,60
Osmolality	296	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 45 min exposure	94	≥ 80 %

Release date 19 April 2017

Released by Vanessa Bierlich

Signature