

Certificate of Analysis GM501 SpermMobil

Batchnumber	SMOBIL-1802
Production date	23 February 2018
Expiry date	31 August 2018
Total volume (mL)	1000

	Results	Specifications
pH	7,24	7,20-7,60
Osmolality	298	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	95	≥ 80 %

Release date 13 March 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann