

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-1805
Production date	27 July 2018
Expiry date	31 January 2019
Total volume (mL)	1354

	Results	Specifications
pH	7,41	7,20-7,60
Osmolality	298	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	86	≥ 80 %

Release date 13 August 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann