

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-1806
Production date	31 August 2018
Expiry date	28 February 2019
Total volume (mL)	1460

	Results	Specifications
pH	7,39	7,20-7,60
Osmolality	299	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	94	≥ 80 %

Release date 17 September 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann