

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-1808
Production date	16 November 2018
Expiry date	31 May 2019
Total volume (mL)	1740

	Results	Specifications
pH	7,27	7,20-7,60
Osmolality	297	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	87	≥ 80 %

Release date 03 December 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann