

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-1907
Production date	11 October 2019
Expiry date	30 April 2020
Total volume (mL)	1990

	Results	Specifications
pH	7,29	7,20-7,60
Osmolality	301	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	95	≥ 80 %

Release date 28 October 2019

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann