

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-1908
Production date	22 November 2019
Expiry date	31 May 2020
Total volume (mL)	1310

	Results	Specifications
pH	7,24	7,20-7,60
Osmolality	297	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	90	≥ 80 %

Release date 10 December 2019

Released by Vanessa Bierlich

Signature