

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-2001
Production date	10 January 2020
Expiry date	31 July 2020
Total volume (mL)	1460

	Results	Specifications
pH	7,25	7,20-7,60
Osmolality	300	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	100	≥ 80 %

Release date 27 January 2020

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann