

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-2005
Production date	19 August 2020
Expiry date	28 February 2021
Total volume (mL)	1580

	Results	Specifications
pH	7,27	7,20-7,60
Osmolality	298	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	91	≥ 80 %

Release date 03 September 2020

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann