

**Certificate of Analysis  
GM501 SpermMobil**

Batchnumber	SMOBIL-2008
Production date	01 December 2020
Expiry date	30 June 2021
Total volume (mL)	1830

	<b>Results</b>	<b>Specifications</b>
pH	7,2	7,20-7,60
Osmolality	298	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	96	≥ 80 %

Release date 16 December 2020

Released by Vanessa Bierlich

Signature