

**Certificate of Analysis  
GM501 SpermMobil**

Batchnumber	SMOBIL-2104
Production date	21 May 2021
Expiry date	30 November 2021
Total volume (mL)	1890

	<b>Results</b>	<b>Specifications</b>
pH	7,38	7,20-7,60
Osmolality	299	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	93	≥ 80 %

Release date 08 June 2021

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*