

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-2108
Production date	21 October 2021
Expiry date	30 April 2022
Total volume (mL)	1890

	Results	Specifications
pH	7,27	7,20-7,60
Osmolality	299	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	91	≥ 80 %

Release date 08 November 2021

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann