

Certificate of Analysis GM501 SpermMobil

Batchnumber	SMOBIL-2110
Production date	29 December 2021
Expiry date	30 June 2022
Total volume (mL)	2390

	Results	Specifications
pH	7,32	7,20-7,60
Osmolality	297	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	94	≥ 80 %

Release date 14 January 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann