

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-2206
Production date	12 October 2022
Expiry date	30 April 2023
Total volume (mL)	2200

	Results	Specifications
pH	7,42	7,20-7,60
Osmolality	299	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	96	≥ 80 %

Release date 27 October 2022

Released by Vanessa Bierlich

Signature