

**Certificate of Analysis  
GM501 SpermMobil**

Batchnumber	SMOBIL-2302
Production date	17 February 2023
Expiry date	31 August 2023
Total volume (mL)	2360

	<b>Results</b>	<b>Specifications</b>
pH	7,31	7,20-7,60
Osmolality	302	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	97	≥ 80 %

Release date 06 March 2023

Released by Nicole Geiger

Signature

