Certificate of Analysis
GM501 SpermMobil

Batchnumber	SMOBIL-2304
Production date	05 May 2023
Expiry date	30 November 2023
Total volume (mL)	2000

	Results	Specifications
pH	7,29	7,20-7,60
Osmolality	301	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	93	≥ 80 %

Release date	23 May 2023
Released by	Nicole Geiger
Signature	pp. N. Geiger

GYNEMED GmbH Co. KG, Lübecker Str. 9, 23738 Lensahn, Germany
phone: +49 4363 90329-0, fax: +49 4363 90329-19, e-mail: info@gynemed.de