

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber SMOBIL-2308
Production date 02 November 2023
Expiry date 31 May 2024
Total volume (ml) 1780

Parameter	Results	Specifications
pH	7,36	7,20-7,60
Osmolality	301	280-320 mOsm/kg
Endotoxin level	< 0,05	< 0,50 EU/mL
Sterility test (according to the current Ph.Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to the test medium)	83	≥ 80 %

Release date 17 November 2023
Released by Nicole Geiger
Signature

pp. N. Geiger

REV 08_00 / 2023-09-15