


**Certificate of Analysis
GM501 SpermMobil**

Batchnumber SMOBIL-2401
 Production date 10 January 2024
 Expiry date 31 July 2024
 Total volume (ml) 2370

Parameter	Results	Specifications
pH	7,44	7,20-7,60
Osmolality	299	280-320 mOsm/kg
Endotoxin level	< 0,05	< 0,50 EU/mL
Sterility test (according to the current Ph.Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to the test medium)	81	≥ 80 %

Release date 26 January 2024
 Released by Nicole Geiger
 Signature



REV 08_00 /2023-09-15