

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber SMOBIL-2406
 Production date 01 August 2024
 Expiry date 28 February 2025
 Total volume (ml) 2170

Parameter	Results	Specifications
pH	7,35	7,20-7,60
Osmolality	298	280-320 mOsm/kg
Endotoxin level	< 0,05	< 0,50 EU/mL
Sterility test (according to the current Ph.Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to the test medium)	88	≥ 80 %

Release date 16 August 2024
 Released by Joana Sturm
 Signature



REV 08_00 / 2023-09-15