

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber SMOBIL-2407
 Production date 06 September 2024
 Expiry date 31 March 2025
 Total volume (ml) 2270

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,41	7,20-7,60
Osmolality	297	280-320 mOsm/kg
Endotoxin level	< 0,05	< 0,50 EU/mL
Sterility test (according to the current Ph.Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to the test medium)	90	≥ 80 %

Release date 23 September 2024
 Released by Nicole Geiger
 Signature 

