

Certificate of Analysis GM501 SpermMobil

Batchnumber	SMOBIL-2606
Production date	03 July 2026
Expiry date	31 January 2027
Total volume (ml)	2290
Product was sterilized by aseptic processing techniques	

Parameter	Results	Specifications
pH at 37°C	7,31	7,20-7,60
Osmolality	287	280-320 mOsm/kg
Endotoxin level	< 0,05	< 0,50 EU/mL
Sterility test (according to the current Ph.Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to the test medium)	86	≥ 80 %

Release date	20 July 2026
Released by	Dr. René Richarz
Signature	

