

**Certificate of Analysis
GM501 SpermStore**

Batchnumber	SPSTORE-1816
Production date	24 July 2018
Expiry date	31 January 2020
Total volume (bottles)	150
Batchnumber HSA	M749A6671

	Results	Specifications
pH	7,51	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	97	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 08 August 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann