

**Certificate of Analysis
GM501 SpermStore**

Batchnumber	SPSTORE-1925
Production date	26 November 2019
Expiry date	31 May 2021
Batchnumber HSA	K917C6871
Total volume (No. of bottles)	455

	Results	Specifications
pH	7,52	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	97	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 12 December 2019

Released by Vanessa Bierlich

Signature