

Certificate of Analysis GM501 SpermStore

Batchnumber	SPSTORE-2011
Production date	29 June 2020
Expiry date	31 December 2021
Batchnumber HSA	B3ALD01343
Total volume (No. of bottles)	620

	Results	Specifications
pH	7,47	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	99	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	14 July 2020
--------------	--------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann