

Certificate of Analysis GM501 SpermStore

Batchnumber	SPSTORE-2112
Production date	07 July 2021
Expiry date	31 January 2023
Batchnumber HSA	B3ALE00033
Total volume (No. of bottles)	1010

	Results	Specifications
pH	7,49	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	97	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	22 July 2021
--------------	--------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann