

**Certificate of Analysis  
GM501 SpermStore**

Batchnumber	SPSTORE-2203
Production date	04 February 2022
Expiry date	31 August 2023
Batchnumber HSA	B3ALF01143
Total volume (No. of bottles)	800

	<b>Results</b>	<b>Specifications</b>
pH	7,43	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	98	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 21 February 2022

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*