

**Certificate of Analysis
GM501 SpermStore**

Batchnumber	SPSTORE-2210
Production date	26 April 2022
Expiry date	31 October 2023
Batchnumber HSA	B3ALF01143
Total volume (No. of bottles)	630

	Results	Specifications
pH	7,52	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 12 May 2022

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann

REV 06_00 / 2019-01-15