

**Certificate of Analysis  
GM501 SpermStore**

Batchnumber	SPSTORE-2224
Production date	27 October 2022
Expiry date	30 April 2024
Batchnumber HSA	B030G002753
Total volume (No. of bottles)	640

	<b>Results</b>	<b>Specifications</b>
pH	7,45	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	96	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 14 November 2022

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*