

**Certificate of Analysis
GM501 SpermStore**

Batchnumber	SPSTORE-2228
Production date	16 December 2022
Expiry date	30 June 2024
Batchnumber HSA	K222B6875
Total volume (No. of bottles)	80

	Results	Specifications
pH	7,47	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	94	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 02 January 2023

Released by Nicole Geiger

Signature

