

**Certificate of Analysis
GM501 SpermStore**

Batchnumber	SPSTORE-2319
Production date	26 July 2023
Expiry date	31 January 2025
Batchnumber HSA	B03H003853
Total volume (No. of bottles)	134

	Results	Specifications
pH	7,49	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	98	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	11 August 2023
--------------	----------------

Released by	Nicole Geiger
-------------	---------------

Signature

pp. N. Geiger