

**Certificate of Analysis  
GM501 SpermStore**

Batchnumber	SPSTORE-2323
Production date	31 August 2023
Expiry date	28 February 2023
Batchnumber HSA	B03H003853
Total number of kits	135

Parameter	Result	Specification
pH	7,51	7,20-7,60
Endotoxin level	< 0,05	<0,25 EU/ml
Concentration HSA	4	3,2-4,8 g/l
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
Sperm Survival Test (4 hours exposure)	97	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	15 September 2023
Released by	Nicole Geiger
Signature	

REV 07\_00 / 2023-09-15