


**Certificate of Analysis  
GM501 SpermStore**

Batchnumber SPSTORE-2415  
 Production date 18 June 2024  
 Expiry date 31 December 2025  
 Batchnumber HSA B03H100413  
 Total number of kits 212

Parameter	Result	Specification
pH	7,44	7,20-7,60
Endotoxin level	< 0,1	<0,25 EU/ml
Concentration HSA	4,0	3,2-4,8 g/l
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	NO GROWTH	No growth
Sperm Survival Test (4 hours exposure)	99	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphillis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 03 July 2024  
 Released by Nicole Geiger  
 Signature



REV 07\_00 / 2023-09-15