

**Certificate of Analysis
GM501 SpermStore**

Batchnumber SPSTORE-2509
 Production date 01 April 2025
 Expiry date 31 October 2026
 Batchnumber HSA B03H121903
 Total number of kits 222
 Product was sterilized by aseptic processing techniques

Parameter	Result	Specification
pH	7,43	7,20-7,60
Endotoxin level	< 0,1	<0,25 EU/ml
Concentration HSA	3,9	3,2-4,8 g/l
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
Sperm Survival Test (4 hours exposure)	98	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 22 April 2025
 Released by Joana Sturm
 Signature



REV 07_00 / 2023-09-15