

## Certificate of Analysis

### GM501 SpermStore

Batchnumber	SPSTORE-2601
Production date	07 January 2026
Expiry date	31 July 2027
Batchnumber HSA	B03J094493
Total number of kits	142

Product was sterilized by aseptic processing techniques

Parameter	Result	Specification
pH	7,43	7,20-7,60
Endotoxin level	< 0,1	<0,25 EU/ml
Concentration HSA	3,7	3,2-4,8 g/l
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
Sperm Survival Test (4 hours exposure)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 22 January 2026

Released by René Richarz

Signature



REV 07\_00 / 2023-09-15