

**Certificate of Analysis
GM501 SpermStore**

Batchnumber SPSTORE-2611
 Production date 08 May 2026
 Expiry date 30 November 2027
 Batchnumber HSA B03J125033
 Total number of kits 17000
 Product was sterilized by aseptic processing techniques

Parameter	Result	Specification
pH	7,43	7,20-7,60
Endotoxin level	< 0,1	<0,25 EU/ml
Concentration HSA	3,7	3,2-4,8 g/l
Sterility test (according to the current Ph. Eur. 2.6.1 / USP <71>)	No growth	No growth
Sperm Survival Test (4 hours exposure)	100	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date 26 May 2026
 Released by Dr. René Richarz
 Signature 

REV 07_00 / 2023-09-15