

Certificate of Analysis GM501 SpermStore

Batchnumber	SPSTORE-2614
Production date	18 June 2026
Expiry date	31 December 2027
Batchnumber HSA	B23K049493
Total number of kits	158
Product was sterilized by aseptic processing techniques	

Parameter	Result	Specification
pH	7,41	7,20-7,60
Endotoxin level	< 0,1	<0,25 EU/ml
Concentration HSA	3,9	3,2-4,8 g/l
Sterility test <small>(according to the current Ph. Eur. 2.6.1 / USP <71>)</small>	No growth	No growth
Sperm Survival Test (4 hours exposure)	98	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	06 July 2026
Released by	Nicole Geiger
Signature	

REV 07_00 / 2023-09-15