

**Certificate of Analysis
GM501 VitriStore Thaw**

Batchnumber	VITRI-1708
Production date	18 July 2017
Expiry date	31 July 2018
Total volume (Kits)	129
Batchnumber HSA	M701A6671

GM501 VitriStore Thaw 1	Results	Specifications
pH	7,27	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	835	805-850 mOsm/kg

GM501 VitriStore Thaw 2	Results	Specifications
pH	7,26	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	551	535-565 mOsm/kg

GM501 VitriStore Thaw 3	Results	Specifications
pH	7,25	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	413	405-435 mOsm/kg

GM501 VitriStore Thaw 4	Results	Specifications
pH	7,28	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	283	270-290 mOsm/kg

Mouse Embryo Assay (1-cell)	90,48	≥ 80 %
-----------------------------	-------	--------

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	03 August 2017
--------------	----------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann